



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/698,125	10/31/2003	Derek Jonathan Harper	P-8609.03 US	8937
27581	7590	08/08/2005	EXAMINER	
MEDTRONIC, INC. 710 MEDTRONIC PARKWAY NE MS-LC340 MINNEAPOLIS, MN 55432-5604			SIRMONS, KEVIN C	
			ART UNIT	PAPER NUMBER
			3763	

DATE MAILED: 08/08/2005

Please find below and/or attached an Office communication concerning this application or proceeding.

## Office Action Summary

Application No.

10/698,125

Applicant(s)

HARPER ET AL.

Examiner

Kevin C. Simons

Art Unit

3763

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

### Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

### Status

- 1) ☒ Responsive to communication(s) filed on 31 October 2003.
- 2a) ☐ This action is FINAL. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

### Disposition of Claims

- 4) ☒ Claim(s) 1-4 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1-4 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

### Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

### Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some \* c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
  - ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

### Attachment(s)

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☐ Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)  
Paper No(s)/Mail Date \_\_\_\_\_
- 4) ☐ Interview Summary (PTO-413)  
Paper No(s)/Mail Date. \_\_\_\_\_
- 5) ☐ Notice of Informal Patent Application (PTO-152)
- 6) ☐ Other: \_\_\_\_\_

## **DETAILED ACTION**

### ***Claim Rejections - 35 USC § 112***

The following is a quotation of the second paragraph of 35 U.S.C. 112:

The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.

Claims 1-4 are rejected under 35 U.S.C. 112, second paragraph, as being indefinite for failing to particularly point out and distinctly claim the subject matter which applicant regards as the invention.

As to claim 1, it is unclear how a hole can have a surface area ranging from about 0.8cm<sup>2</sup> to about 5.0cm<sup>2</sup>. It would seem impossible for a hole, which has an opening to have a surface area.

### ***Claim Rejections - 35 USC § 103***

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 1-4 are rejected under 35 U.S.C. 103(a) as being unpatentable over Goldberg et al U.S. Pat. No. 5,102,404.

Goldberg discloses a drip chamber in a cerebral spinal fluid drainage system comprising: a tube having an outer surface (figs. 2 and 4); a vent (fig. 4); a filter (36) made of hydrophobic porous material wherein the pore size of the filter ranges from greater than .45um to about 5.0um (36). However, it is not clear if Goldberg discloses a

vent having a surface area ranging from about 0.08cm<sup>2</sup> to about 5.0cm<sup>2</sup> primarily because the claim is unclear. However, it would have been obvious matter of mechanical efficiency with one have ordinary skill in the art at the time the invention was made to vary the size of the surface area ranging from about 0.08cm<sup>2</sup> to about 5.0cm<sup>2</sup>, since applicant indicated in the specification that the aforementioned size range is preferable, but not critical, and since such a modification would have involved a mere change in the size of a component. A change in size is generally recognized as being within the level of ordinary skill in the art. *In re Rose*, 105 USPQ 237 (CCPA 1955). Basically, applicant's ranges lacks criticality and One of ordinary skill in the art, furthermore, would have expected Applicant's invention to perform equally well with the surface area of Goldberg because they are in the same field of endeavor. Therefore, it would have been an obvious matter of design choice to modify a drip chamber of Goldberg to obtain the invention as specified in claim 1.

As to claims 2-4, (figs. 1 and 2).

### ***Double Patenting***

The nonstatutory double patenting rejection is based on a judicially created doctrine grounded in public policy (a policy reflected in the statute) so as to prevent the unjustified or improper timewise extension of the "right to exclude" granted by a patent and to prevent possible harassment by multiple assignees. See *In re Goodman*, 11 F.3d 1046, 29 USPQ2d 2010 (Fed. Cir. 1993); *In re Longi*, 759 F.2d 887, 225 USPQ 645 (Fed. Cir. 1985); *In re Van Ornum*, 686 F.2d 937, 214 USPQ 761 (CCPA 1982); *In re Vogel*, 422 F.2d 438, 164 USPQ 619 (CCPA 1970); and, *In re Thorington*, 418 F.2d 528, 163 USPQ 644 (CCPA 1969).

A timely filed terminal disclaimer in compliance with 37 CFR 1.321(c) may be used to overcome an actual or provisional rejection based on a nonstatutory double patenting ground provided the conflicting application or patent is shown to be commonly owned with this application. See 37 CFR 1.130(b).

Effective January 1, 1994, a registered attorney or agent of record may sign a terminal disclaimer. A terminal disclaimer signed by the assignee must fully comply with 37 CFR 3.73(b).

Claims 1-4 are provisionally rejected under the judicially created doctrine of obviousness-type double patenting as being unpatentable over claims 1-4, 6-13, 16-22, 24-31, 33-37, 39 and 40 of copending Application No. 09/266,674. Although the conflicting claims are not identical, they are not patentably distinct from each other because the claims in both applications are directed to a drip chamber comprising: a tube, vent, filter and surface area.

This is a provisional obviousness-type double patenting rejection because the conflicting claims have not in fact been patented.

### ***Conclusion***

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Kevin C. Sirmons whose telephone number is 571-272-4965. The examiner can normally be reached on Monday-Friday 6:30-4:00 ALT FRI.

The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Art Unit: 3763

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

Kevin C. Sirmons  
Primary Examiner  
Art Unit 3763  
8/4/05

A handwritten signature in cursive script that reads "Kevin C. Sirmons".